



785 NE Dixie Hwy.
Jensen Beach, FL 34957
Telephone: (772) 225-6932
FAX: (772) 225-6291
www.koppdevelopment.com

CERTIFICATION

December 22, 2010

To whom it may concern:

We have tested the MRI Non-Ferromagnetic 20" Wheelchair sold by MRIequip.com both with our FerrAlert™ HALO ENTRY and with our FerrAlert™ HALO PRESCREEN. The referenced wheelchair did not trigger an alarm with either product.

We can therefore certify that the MRI Non-Ferromagnetic 20" Wheelchair sold by MRIequip.com does not cause extraneous alarms when used unmodified with our ferromagnetic detectors

Sincerely,

A handwritten signature in black ink, appearing to read "Keith Kopp", written in a cursive style. The signature is positioned above the printed name and title.

Keith Kopp
President